

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 114年 12月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2025/12/1	4.1	5.5	4	5	3.2	4.1	-20.3	-16.6	-79.6	-77.1	19.4	19.7
2025/12/2	4.1	4.8	4	4.7	3.2	4.1	-19.7	-17.9	-78.3	-77.1	19.3	20.2
2025/12/3	4.1	4.8	4	4.5	3.1	3.9	-20.8	-15.6	-78.4	-77.2	19.5	20.2
2025/12/4	3.9	5.1	3.8	4.4	3.1	4.3	-20.1	-17.1	-78.4	-77.2	19.4	20.4
2025/12/5	3.9	5.1	4	4.5	3.2	3.9	-19.9	-17.8	-78.4	-77.2	19.7	20.3
2025/12/6	4.3	4.7	4	4.5	3.2	3.9	-19.1	-18.3	-78.3	-77.2	19.7	19.8
2025/12/7	4.3	4.7	4	4.5	3.2	3.9	-18.9	-18.8	-78.3	-77.2	19.7	19.8
2025/12/8	4.2	6.1	3.9	4.5	2.8	4.2	-19.3	-18.4	-78.4	-77.2	19.7	20.4
2025/12/9	4.1	5.1	3.8	4.6	2.7	4	-19.6	-18.7	-78.4	-77.2	19.8	20.5
2025/12/10	4.1	5.3	4	4.6	2.8	4.6	-19.6	-18.3	-78.4	-77.2	19.9	20.6
2025/12/11	4	5.7	4	4.8	3.9	4.7	-19.8	-17.3	-78.5	-77.2	19.9	20.5
2025/12/12	4	5.6	4	5	4	4.4	-19.8	-18.7	-78.4	-77.2	19.8	20.4
2025/12/13	4.2	4.8	4.2	4.8	4.1	4.3	-19.1	-18.8	-78.4	-77.1	19.6	19.9
2025/12/14	4.2	4.8	4.2	4.8	4.1	4.3	-19.1	-19	-78.3	-77.2	19.1	19.6
2025/12/15	4.2	5.5	4.2	4.8	4	4.3	-19.9	-18.4	-79.3	-77.1	19	20
2025/12/16	3.6	4.8	4.1	5.2	3.9	4.6	-19.6	-18.9	-79.3	-77.7	18.7	20
2025/12/17	3.7	4.5	4.1	4.7	3.9	4.4	-19.9	-19.1	-79.3	-77.7	18.2	19.3
2025/12/18	3.6	4.8	4.1	4.7	3.9	4.3	-20.8	-19.1	-79.3	-77.7	18.2	19.3
2025/12/19	3.5	4.6	4	4.8	3.3	4.3	-19.8	-19.4	-79.3	-77.7	18.8	19.5
2025/12/20	3.7	4.5	4.1	4.6	3.3	3.7	-19.6	-19.6	-79.3	-77.7	18.8	19
2025/12/21	3.7	4.5	4.1	4.6	3.3	3.7	-19.6	-19.6	-79.2	-77.7	18.4	18.8
2025/12/22	3.6	4.9	4	4.7	3.4	4.4	-19.6	-18.7	-79.8	-77.7	18.3	19
2025/12/23	3.3	4.6	3.6	4.6	3.8	4.4	-19.1	-18.3	-79.3	-77.7	18.4	19.1
2025/12/24	3.3	5.2	3.9	4.5	3.8	4.3	-19.1	-18.7	-79.3	-77.7	18.6	19.2
2025/12/25	3.9	4.7	3.9	4.4	3.9	4.3	-19.1	-18.9	-79.3	-77.7	18.2	18.8
2025/12/26	3.7	4.7	3.9	4.7	3.9	4.4	-19.7	-18.8	-79.3	-77.7	18.2	19
2025/12/27	3.9	4.6	4	4.5	3.9	4.3	-18.9	-18.7	-79.3	-77.7	18.2	18.4
2025/12/28	3.9	4.6	4	4.5	3.9	4.3	-18.8	-18.8	-79.3	-77.7	18.2	18.3
2025/12/29	3.7	4.8	3.9	4.5	3.2	5.5	-19.3	-17.8	-79.3	-77.7	18.2	18.8
2025/12/30	3.7	4.7	3.2	5.9	3.2	3.9	-18.9	-18.7	-79.3	-77.7	18.3	18.9
2025/12/31	3.4	4.4	3.2	3.8	3.2	3.9	-18.8	-18	-79.3	-77.7	18.4	19

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L
 註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 邱保祥
 師(三)級藥師 115
 0102

總藥師:

藥學部 溫雅如
 師(二)級藥師 115
 0102

業管主管:

藥學部 李婉詩
 師(二)級藥師 115
 0102