

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 114 年 12 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		2~8℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2025/12/1	4.4	4.9	4.4	5.1	4.2	5.4	5.9	6.2	4.4	5.3	19.9	20.3
2025/12/2	4	4.5	4.4	5.5	4.2	4.7	5.9	6.1	4.7	5.7	20.1	20.9
2025/12/3	4	4.3	4.3	5.3	4.1	4.9	5.8	6.2	4.4	5.8	20.8	21.1
2025/12/4	4.1	4.5	4.7	6.2	4	4.6	3.9	6.1	4.2	5.1	20.7	21.1
2025/12/5	4.1	4.3	4.7	5.4	4	4.7	3.9	4.9	4.4	5.2	20.9	21.3
2025/12/6	4.1	4.3	4.7	5.1	4.2	4.7	3.9	4.7	4.4	4.8	21.2	21.3
2025/12/7	4.1	4.3	4.7	5.1	4.2	4.7	3.9	4.7	4.3	4.8	21.1	21.2
2025/12/8	4.1	4.4	4.7	6.4	4.2	4.8	3.9	4.7	4.1	4.8	21.1	21.4
2025/12/9	4.1	5.4	4.8	5.4	4.1	4.7	3.9	4.8	3.9	4.7	21.3	21.7
2025/12/10	4.9	5.2	4.4	5.3	4.2	4.8	3.9	4.8	3.9	4.7	21.4	21.8
2025/12/11	4.3	5.7	4.3	4.8	4.2	4.8	4	4.8	3.6	4.8	21.5	21.8
2025/12/12	4.3	4.4	4.4	5.2	4.2	5.3	3.9	4.7	3.7	4.8	21.3	21.7
2025/12/13	4.3	4.4	4.7	5.1	4.3	4.8	4	4.7	3.7	4.8	21.4	21.7
2025/12/14	4.3	4.4	4.7	5.1	4.2	4.7	4	4.7	3.7	4.8	20.7	21.4
2025/12/15	4.3	4.6	4.6	5.4	4.2	5.8	4	4.8	3.8	5.3	20.5	20.8
2025/12/16	4.3	5.1	4.5	5.4	4.2	5.2	3.8	4.7	3.8	5.2	20.4	21.2
2025/12/17	4.4	5.4	4.4	5.6	4.3	5.1	3.7	4.7	3.9	4.7	19.9	20.6
2025/12/18	4.4	4.7	4.4	5.9	4.4	5.1	3.8	4.6	3.9	5.2	19.9	20.6
2025/12/19	4.1	4.6	4.7	5.1	4.6	5.1	3.7	4.6	4	4.8	20.4	20.8
2025/12/20	4.2	4.3	4.7	5.1	4.6	5.1	3.7	4.7	4	4.8	20.4	20.6
2025/12/21	4.2	4.3	4.7	5.1	4.6	5.1	3.7	4.7	4	4.9	19.9	20.4
2025/12/22	4.2	4.4	4.7	5.2	4.6	5.3	3.7	4.7	3.9	5.6	19.8	20.2
2025/12/23	4.2	4.4	4.7	5.2	4.5	5.1	3.7	4.7	3.9	4.9	19.9	20.3
2025/12/24	4.2	4.6	4	5.6	4.5	5.1	3.9	4.8	3.9	4.8	20.1	20.5
2025/12/25	4.2	4.4	4.3	4.7	4.5	5.1	3.9	4.8	3.9	4.9	19.8	20.3
2025/12/26	4.2	4.4	4.2	4.8	4.5	5.1	3.9	4.8	4	4.8	19.8	19.9
2025/12/27	4.3	4.4	4.3	4.7	4.5	4.9	3.9	4.8	4.1	4.8	19.7	19.8
2025/12/28	4.3	4.4	4.3	4.7	4.5	4.9	3.9	4.8	4.1	4.8	19.6	19.7
2025/12/29	4.2	4.7	4.3	4.8	4.3	5.6	3.9	4.8	4	4.8	19.5	19.9
2025/12/30	4.3	5.3	4.3	4.7	4.3	5.1	3.9	4.7	3.8	5.6	19.7	20.1
2025/12/31	4.6	5.3	4.2	4.7	4.3	5.2	3.9	4.7	3.8	4.7	19.9	20.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 邱保祥 115 總藥師: 0102
藥學部 溫雅如 115 師(三)級藥師 0102

業管主管:

藥學部 李婉詩 0106
師(二)級藥師 代