

STATEMENT OF ACCURACY

ILog Temperature Data loggers

Product Code: **EL-IN-D-32-L**
Serial Number: **2052-0129**

Date: **29 Jan 2021**



Crypapak
The Perfect Degree of Protection

Crypapak Verification Technologies, Inc.

72 Executive Circle
Roanoke, VA 24012
Phone: 540-254-1433 | Fax: 540-254-2433

techsupport@crypapak.com

www.crypapak.com

ILog data loggers are calibrated for temperature in CTT's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST). Temperature Reference Meter: Fluke -Hart Scientific 1521 / 1522. Digital Thermometer Readout w/Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.

Temp. Code	Full Range	Temp. Range	Accuracy	Temp. Range	Accuracy	Temp. Range	Accuracy
X	C	-20°C to +100°C	-20°C to 0°C	± 0.30°C	0°C to +70°C	± 0.25°C	+70°C to +100°C ± 0.30°C
✓	D	-40°C to +70°C	-40°C to 0°C	± 0.35°C	0°C to +70°C	± 0.25°C	
X	N	-100°C to +40°C	-100°C to -80°C	± 1.0°C	-80°C to -50°C	± 0.50°C	-50°C to +40°C ± 0.30°C
X	Z	0°C to +150°C	0°C to +90°C	± 0.30°C	+90°C to +120°C	± 0.45°C	+120°C to +150°C ± 0.75°C

Penny Williams
Penny Williams (Calibration Tech.)

Parminder Singh
Parminder Singh (Quality Manager)

Authorized Quality Document for Crypapak Verification Technologies. Reference: ISO 9001:2015, ACC-031 Revision 5 Page Form Page 1/1.

ISO 9001:2015
Certified Company