

STATEMENT OF ACCURACY

ILog Temperature Data loggers

Product Code: **EI-IN-D-32-L**
Serial Number: **2123-0026**

Date: **22 Jun 2021**



Crypapak
The Perfect Degree of Protection

Crypapak Verification Technologies, Inc.
72 Executive Circle
Roanoke, VA 24012
Phone: 540-254-1433 | Fax: 540-254-2433
TechSupport@crypapak.com
www.crypapak.com

ILog data loggers are calibrated for temperature in CPT's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST). Temperature Reference Meter: Fluke-Hart Scientific 1521 / 1522. Digital Thermometer Readout w. Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.

Temp. Code	Full Range	Temp. Range	Accuracy	Temp. Range	Accuracy	Temp. Range	Accuracy	
X	C	-20°C to +100°C	-20°C to 0°C	± 0.30°C	0°C to +70°C	± 0.25°C	+70°C to +100°C	± 0.30°C
✓	D	-40°C to +70°C	-40°C to 0°C	± 0.35°C	0°C to +70°C	± 0.25°C		
X	N	-100°C to +40°C	-100°C to -80°C	± 1.0°C	-80°C to -50°C	± 0.50°C	-50°C to +40°C	± 0.30°C
X	Z	0°C to +150°C	0°C to +90°C	± 0.30°C	+90°C to +120°C	± 0.45°C	+120°C to +150°C	± 0.75°C

Perry Williams
Perry Williams (Calibration Tech.)

Parminder Singh
Parminder Singh (Quality Manager)

ISO 9001:2015
Certified Company

Authorized Quality Document for Crypapak Verification Technologies Reference: ISO 9001:2015, ACCT-0131 Revision 5 Page Form Page 1 of 1.