

STATEMENT OF ACCURACY

iLog Temperature Data loggers

Product Code: **EI-IN-D-32-L**

Serial Number: **2123-0025**

Date: **22 Jun 2021**



Cryopak
The Perfect Degree of Protection

Cryopak Verification Technologies, Inc.
72 Executive Circle
Roanoke, VA 24012
Phone: 540-254-1433 | Fax: 540-254-2433
techsupport@cryopak.com
www.cryopak.com

iLog data loggers are calibrated for temperature in CVT's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST). Temperature Reference Meter: Fluke -Hart Scientific 1521 / 1522 Digital Thermometer Readout w/ Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.

Temp. Code	Full Range	Temp. Range	Accuracy	Temp. Range	Accuracy	Temp. Range	Accuracy
×	C	-20°C to +100°C	-20°C to 0°C	± 0.30°C	0°C to +70°C	± 0.25°C	+70°C to +100°C ± 0.30°C
✓	D	-40°C to +70°C	-40°C to 0°C	± 0.35°C	0°C to +70°C	± 0.25°C	
×	N	-100°C to +40°C	-100°C to -80°C	± 1.0°C	-80°C to -50°C	± 0.50°C	-50°C to +40°C ± 0.30°C
×	Z	0°C to +150°C	0°C to +90°C	± 0.30°C	+90°C to +120°C	± 0.45°C	+120°C to +150°C ± 0.75°C

Penny Williams
Penny Williams (Calibration Tech.)

Parminder Singh
Parminder Singh (Quality Manager)

ISO 9001:2015
Certified Company

Authorized Quality Document for Cryopak Verification Technologies. Reference: ISO 9001:2015. ACC-031 Revision 5 Page Form Page 1 1.