

STATEMENT OF ACCURACY

iLog Temperature Data loggers

iLog data loggers are calibrated for temperature in CVT's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST). Temperature Reference Meter: Fluke -Hart Scientific 1521 / 1522 Digital Thermometer Readout w/ Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.

Product Code: **EI-IN-D-32-L**

Serial Number: **2123-0028**

Date: **22 Jun 2021**



Cryopak
The Perfect Degree of Protection

Cryopak Verification Technologies, Inc.
72 Executive Circle
Roanoke, VA 24012
Phone: 540-254-1433 | Fax: 540-254-2433
techsupport@cryopak.com
www.cryopak.com

	Temp. Code	Full Range	Temp. Range	Accuracy	Temp. Range	Accuracy	Temp. Range	Accuracy
×	C	-20°C to +100°C	-20°C to 0°C	± 0.30°C	0°C to +70°C	± 0.25°C	+70°C to +100°C	± 0.30°C
✓	D	-40°C to +70°C	-40°C to 0°C	± 0.35°C	0°C to +70°C	± 0.25°C		
×	N	-100°C to +40°C	-100°C to -80°C	± 1.0°C	-80°C to -50°C	± 0.50°C	-50°C to +40°C	± 0.30°C
×	Z	0°C to +150°C	0°C to +90°C	± 0.30°C	+90°C to +120°C	± 0.45°C	+120°C to +150°C	± 0.75°C

Penny Williams
Penny Williams (Calibration Tech.)

Parminder Singh
Parminder Singh (Quality Manager)

ISO 9001:2015
Certified Company

Authorized Quality Document for Cryopak Verification Technologies. Reference: ISO 9001:2015. ACC-031 Revision 5 Page Form Page 1 of 1.